

### Semi fowler Bed, Hospital

S.N.	Purchaser's Specification (FY:82/83)	Bidder's Compliance Sheet
	<b>Bed, Hospital</b>	
	<b>Manufacturer</b>	
	<b>Brand</b>	
	<b>Type/Model</b>	
	<b>Country of Origin</b>	
<b>1</b>	<b>Description of Function</b>	
1.1	A hospital bed is a bed specially designed for hospitalized patients in need of patient ease. These beds have special features both for the comfort and well-being of the patient and for the convenience of hospital staff.	
<b>2</b>	<b>Operational Requirements</b>	
2.1	The patient bed shall be made of solid steel construction with anti-corrosive and antirust treated baked hard epoxy powder coating.ABS or steel made side panel.	
<b>3</b>	<b>System Configuration</b>	
3.1	Hospital Bed epoxy powder coated with panels and mattress	
<b>4</b>	<b>Technical Specifications</b>	
4.1	Bed base shall be anti-corrosive and antirust treated epoxy powder coated welded steel bar or epoxy powder coated 18G perforated sheet top to improve ventilation.	
4.2	The patient bed shall be fixed height where the backrest section could be elevated by mechanical hand crank located at the foot-end of the bed.	
4.3	It shall come with one dual hook anti-corrosive and antirust treated epoxy powder coated or 304 grade stainless steel IV rod.	
4.5	All 4 legs of the hospital bed shall be capped with heavy duty rubber footings or Noiseless castors wheel with brakes(2)	
4.6	It shall have ABS or SS Head & Foot Panels	
4.7	Overall approximate dimension: not less than 2000 mm length, 93cm width, 68cm height	
<b>5</b>	<b>System Accessories, spares and consumables</b>	
5.1	Two sectional mattress,high quality,must be included in the offer	
5.2	All standard accessories/consumables/parts required for the proper operation of the above item shall be included in the offer.	
<b>6</b>	<b>Operating Environment</b>	
6.1	The system offered shall be designed to store and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.	
<b>7</b>	<b>Standards and Safety Requirements</b>	
7.1	Must submit ISO 9001 or ISO 13485:2003/AC:2007	
<b>8</b>	<b>User Training</b>	

  
 Medical Engineer  
 NEC No. 365 "A"

<b>S.N.</b>	<b>Purchaser's Specification (FY:82/83)</b>	<b>Bidder's Compliance Sheet</b>
8.1	Not applicable.	
<b>9</b>	<b>Warranty</b>	
9.1	Warranty for 2 years after acceptance.	
<b>10</b>	<b>Maintenance Service During Warranty Period</b>	
10.1	Standard warranty conditions are applicable.	
<b>11</b>	<b>Installation and Commissioning</b>	
11.1	Must supply preassembled unit, ready to use.	
<b>12</b>	<b>Documentation</b>	
12.1	Users/Instructions manual shall be provided in English.	

  
 asst-Medical Engineer  
 NEC No. 365 "A"